



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

09/720,762

12/28/2000

Kazuyuki Yanase

114474-14-FESI00002US

5214

38492

7590

10/17/2008

WILLKIE FARR & GALLAGHER LLP
INTELLECTUAL PROPERTY LEGAL ASSISTANTS
787 SEVENTH AVE
NEW YORK, NY 10019-6099

EXAMINER

HAMO, PATRICK

ART UNIT

PAPER NUMBER

3746

MAIL DATE

DELIVERY MODE

10/17/2008

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No. 09/720,762	Applicant(s) YANASE ET AL.	
	Examiner PATRICK HAMO	Art Unit 3746	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 05 August 2008.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1,3,6-9,11 and 13-40 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1,3,6,9-11 and 13-40 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Continued Examination Under 37 CFR 1.114

A request for continued examination under 37 CFR 1.114, including the fee set forth in 37 CFR 1.17(e), was filed in this application after final rejection. Since this application is eligible for continued examination under 37 CFR 1.114, and the fee set forth in 37 CFR 1.17(e) has been timely paid, the finality of the previous Office action has been withdrawn pursuant to 37 CFR 1.114. Applicant's submission filed on July 10, 2008 has been entered.

Claim Objections

Claims 1, 9, 23 and 33 are objected to because of the following informalities: the use of single brackets is not a proper form of indicating deleted subject matter. To delete subject matter consisting of more than five characters, strike-through is to be used. In the event that brackets are used to indicate deleted subject matter, it is required that they be double brackets ([[example]]). However, brackets would not be suitable in the instant case because the deletion was more than five characters. See 37 CFR 1.121(b)(1)(ii). Appropriate correction is required.

Claim Rejections - 35 USC § 103

The text of those sections of Title 35, U.S. Code not included in this action can be found in a prior Office action.

Art Unit: 3746

Claims 1, 3, 6-9, 16-28 and 30-32 are rejected under 35 U.S.C. 103(a) as being unpatentable over Trull et al., 6,080,136 in view of Sudo et al., 5,009,646.

Trull, in figure 6, disclose a syringe gasket (70) wherein a peripheral side surface (80) of the gasket is in contact with an inner surface of the syringe barrel (60). A restriction and a tapered slant with a first and second plunger diameter (labeled by the examiner on Figure 6 attached to the end of this office action) is provided, and a periphery of a bottom surface of the gasket that is not in contact with the liquid is formed into a tapered slant (also labeled by the examiner on Figure 6).

Trull does not teach that one or both of the peripheral side surfaces that is in contact with an inner surface of the syringe barrel and a surface of the gasket that is in contact with the liquid is laminated with polyethylene fluoride resin.

However, Sudo teaches a syringe gasket (2) coated (3) with a thermoplastic resin such as polyethylene (col. 2 lines 10-24) to provide sealing and lubrication while preventing contamination from liquid lubricants (col. 1 lines 48-55). In regard to the claimed limitation that the restriction has an inner diameter ratio of 0.8 or more and less than 1.0 of a diameter of the gasket, this constitutes a change in proportion of the gap 6 taught by Sudo. In *Gardner v. TEC Systems, Inc.*, 725 F.2d 1338, 220 USPQ 777 (Fed. Cir. 1984), cert. denied, 469 U.S. 830, 225 USPQ 232 (1984), the Federal Circuit held that, where the only difference between the prior art and the claims was a recitation of relative dimensions of the claimed device and a device having the claimed relative dimensions would not perform differently than the prior art device, the claimed device was not patentably distinct from the prior art device.

Art Unit: 3746

Therefore it would have been obvious to one of ordinary skill in the art at the time the invention was made to modify the Trull et al. gasket by using a gasket material of polyethylene, as taught by Sudo et al., in order to eliminate the need for a lubricant and serve as a protective sealant coating for the gasket.

With respect to the fact that Sudo teaches the coating for more than just the peripheral side surface, namely the top surface in contact with the pumped liquid as a sealant, it is brought to applicant's attention that the bottom surface is not coated. This surface is not coated because coating it with a lubricant or sealant is not required. The coating has two functions, namely to lubricate and to seal. For the side surface, both of these functions are utilized. For the top portion, only the function of sealant is being utilized. In the present application, the coating and its function (sealing) are removed from the top surface. This does not add patentable weight because it would have been obvious to omit the coating where the function attributed to the coating is not desired or required. See MPEP §2144.04(2)(a).

Furthermore, Trull does not teach that the first and second diameter of the tapered slant have a difference between about .5mm and about 5mm or that the gasket's inner diameter, its height, its first diameter, or its second diameter. With respect to the specified gasket dimensions in the claims 3, 6-8, 19, 21,22, and 24-28, where the general conditions of a claim are disclosed in the prior art, it is not inventive to discover the optimum or workable ranges by routine experimentation. In re Swain et al., 33 CCPA (Patents) 1250, 156 F.2d 239, 70 USPQ 412; Minnesota Mining and Mfg. Co. v.

Art Unit: 3746

Coe, 69 App. D.C. 217, 99 F.2d 986, 38 USPQ 213; Allen et al. v. Coe, 77 App. D.C. 324, 135 F.2d 11, 57 USPQ 136.

With respect to claim 9, a second tapered (see examiner's marked up Figure 6) slant is formed between the peripheral side surface of the gasket (70) that is in contact with an inner surface of the syringe barrel (60) and the restriction (see examiner's marked up Figure 6). The gasket tightly closing the liquid is an obvious requirement for the syringe to pump fluid properly. A recitation with respect to the material intended to be worked upon by a claimed apparatus, in this case a contrast medium, does not impose any structural limitations upon the claimed apparatus, which differentiates it from the prior art apparatus satisfying the structural limitations of the claims, as is the case here.

Claims 14, and 15 are rejected under 35 U.S.C. 103(a) as being unpatentable over the references as applied to claims 1 and 9 above, and further in view of Akaike et al., 5,061,247.

The references as applied to claims 1 and 9 above teach all of the limitations substantially as claimed except for the following: the gasket being made integrally of a material with JIS hardness of 55 to 60.

However, Akaike et al., in column 5 lines 58-59, disclose, that a hardness of JIS of 20-85 is optimal for gaskets applied to syringe devices. This general range covers the applicants claimed range. Therefore it would have been obvious to one of ordinary skill in the art at the time the invention was made to further modify references as applied to

Art Unit: 3746

claims 1 and 9 above with a gasket material of 20-85 JIS hardness in order to achieve optimum gasket functionality within a syringe device. With respect to the specific range of JIS hardness 55 to 60, where the general conditions of a claim are disclosed in the prior art, it is not inventive to discover the optimum or workable ranges by routine experimentation. In re Swain et al., 33 CCPA (Patents) 1250, 156 F.2d 239, 70 USPQ 412; Minnesota Mining and Mfg. Co. v. Coe, 69 App. D.C. 217, 99 F.2d 986, 38 USPQ 213; Allen et al. v. Coe, 77 App. D.C. 324, 135 F.2d 11, 57 USPQ 136. Other than the range being preferred for presumably general optimum device function, the applicant has provided no criticality or unexpected or non-obvious advantage over choosing one this particular range. The coating taught by the prior art is integral in that the coating and the plunger coated form an integral plunger assembly.

Claims 11 and 13 and 29 are rejected under 35 U.S.C. 103(a) as being unpatentable over the references as applied to claims 1 and 9 above, and further in view of Higashikawa, 5,830,193.

The references as applied to claim 1 above teach all of the limitations substantially as claimed except for the following: that the syringe includes a luer lock.

However, Higashikawa in Figure 1a-1c, 7a, and 7b, teaches that luer lock mechanisms (37, 30, 22) have been especially common in medical syringes (21) for mounting needles (32) - see column 7 line 49.

Therefore it would have been obvious to one of ordinary skill in the art at the time the invention was made to further modify the references as applied to claims 1 and 9

Art Unit: 3746

above by incorporating the luer locking mechanism, as taught by Higashikawa, in order to allow for needle mounting.

Claims 33-40 are rejected under 35 U.S.C. 103(a) as being unpatentable over the references as applied to claim 1 in view of Vacca, 5,531,255.

The references as applied to claim 1 above teach all of the limitations substantially as claimed except for the following: that only wherein the peripheral side surface that contacts the inner surface of the syringe barrel is laminated with silicon.

However, Vacca teaches that adding lubricant to a syringe improves its performance and that silicon is a suitable lubricant (col. 3, ll. 14-18).

Therefore, it would have been obvious to one having ordinary skill in the art at the time the invention was made to have modified the references if claim 1 above with the silicon lubrication of Vacca in order to improve the performance of the syringe.

Response to Arguments

Applicant's arguments filed July 10, 2008 have been fully considered but they are not persuasive.

Applicant argues that the limitation that the restriction has an inner diameter with a ratio of 0.8 or more and less than 1.0 to a diameter of the gasket patentably distinguishes over the art of record. However, as indicated in the rejection under 35 U.S.C. 103 above, this limitation was an obvious change in size in view of the combination of Sudo and Trull.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to PATRICK HAMO whose telephone number is (571)272-3492. The examiner can normally be reached on M-F 8:30-5.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Devon Kramer can be reached on 571-272-7118. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Devon C Kramer/
Supervisory Patent Examiner, Art
Unit 3746

/Patrick Hamo/
Patent Examiner, AU 3746